

检验报告

Certificate of Analysis

产品信息 Product Information			
生产厂家 Vendor	武汉华大智造科技有限公司 Wuhan MGI Tech Co.,Ltd.	产品类型 Product Type	成品 Finished Goods
产品货号 Part No.	940-000422-00	产品名称 Product Name	测序反应通用试剂盒（联合探针锚定聚合测序法） Universal Sequencing Reation Kit
规格 Spec	1测试/盒	批次 Lot No.	WLC24019
失效日期 Expiration date	2025.07.06	储存条件 Storage Condition	-25℃~-15℃
主要成分 Main Component	测序载片、DNB制备试剂、DNB加载试剂和测序试剂 Sequencing chip, DNB preparation reagent, DNB loading reagent and sequencing reagent		
报告内容 Report Contents			
检验项目 Inspection Content	标准 Specification	检验结果 Results	
1.非性能 Non-Performance	1) 内外包装应完好无破损，标签字迹清晰无污渍。产品名称、生产企业、批号及效期正确无误。The inner and outer package should be complete and undamaged, label free from stains and legible. The product name, manufacturer, batch number and validity period are correct.	符合Conforms	
	2) 各组分信息、数量正确。 The information and quantity should be correct.	符合Conforms	
	3) 液体试剂无沉淀及絮状物，无肉眼可见颗粒，无漏液。磁珠溶液为棕色悬浮状液体。Liquid reagent has no precipitation and flocculation, no visible particles, and no liquid leakage.The beads reagent is brown suspended liquid.	符合Conforms	
	4) 装量符合要求。 The volume of reagents should meet the standard of quality inspection.	符合Conforms	
2.性能 Performance	关键性能测试结果满足如下标准： The result of key performance test should meet the quality standard as follows: 准确率≥99.0% Accuracy≥99.0%	符合Conforms	
	重复性：CV值不大于5%（n=5） Repeatability: CV≤5%（n=5）	符合Conforms	
报告结论 Conclusion	合格PASS 合格 合格		

检验人/日期: 黄建华 2024/11/18
Inspector/Date:

批准/日期: [Red Stamp: 华大智造科技有限公司] 2024/11/18
Approved by/Date: